

Teledyne FLIR LLC

10F, No. 57, Zhouzi Street, NeiHu District, Taipei city, 114, Taiwan

Phone: +886-2-26560399 / Fax: +886-2-26578016

Declaration of Conformity

Extech Model: RH250W
Description: Hygro-Thermometer
Date of Issue: 31-Aug-21

We, Teledyne FLIR LLC., 10F, No. 57, Zhouzi Street, NeiHu District, Taipei city, 114, Taiwan declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2014/30/EU
Report Number: ED210603093E; ES210603052W1
Report Date of Issue: Jun-19, 2021; Jun-28, 2021

Standards:

EN 61326-1:2013; EN 61326-2-3:2013
EN 301 489-1 V2.2.3:2019; EN 301 489-17 V3.2.4:2020

LVD Directive: 2014/35/EU
Report Number: ED210603096S; ES210603052W3
Report Date of Issue: Jun-25, 2021; Jun-28, 2021

Standards:

EN 61010-1:2010 + A1:2019
EN 62479:2010; EN 50663:2017

RED Directive: 2014/53/EU
Report Number: ES210603052W2
Report Date of Issue: Jun-28, 2021

Standards:

EN 300 328 V2.2.2:2019

RoHS Directive: EU Directive 2015/863/EU
REACH Directive: EC/1907/2006
WEEE Directive Compliance

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive, EMC Directive, RED Directive, RoHS Directive, REACH and WEEE for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive, EMC Directive, RED Directive, RoHS Directive, REACH and WEEE.



Hank Tsai / QA Manager